- the outer layer is made of chlorine-free bleached cotton gauze
- the inner layer is made of absorbent cotton wool
- soft and absorbent
- gauze edges folded inside the pad, do not fray
- compact structure of the absorbent core gives even pressure on the eyeball

Application:

- for use during and after ophthalmological procedures

Types:

- sterile

Size and packing:

- available sizes: 5 cm x 6 cm
- sterile: packed by 1 pc

NOTE: For invasive procedures use the product only in an unwound state.



MD

Sterile devices are certified by the Notified Body TÜV Nord Polska Sp. z o.o. and bear the CE mark with number 2274

>>

©Copyright 2026 en.matopat-global.com | Last updated: 04.11.2026